

Department of Forensic Science  
Breath Alcohol Section  
Instrument Location History  
1/1/1990 To 4/9/2015

JMS  
Jed

---

Instrument Serial Number: 010620 As of 09-Apr-15

---

Start Date	End Date	Agency Name
08-Apr-15		DFS Central Lab
22-Oct-14	08-Apr-15	Buena Vista PD
09-Oct-14	22-Oct-14	DFS Central Lab
08-May-14	09-Oct-14	Page County SO
30-Apr-14	08-May-14	DFS Central Lab
08-Jan-09	30-Apr-14	Louisa County SO
15-Aug-08	08-Jan-09	DFS Central Lab

**INTOX EC/IR II**  
**Quality Assurance Worksheet**

*AMS*  
*Feb*

Instrument Serial Number **010620** Worksheet Start Date **4/8/2015**  
Location **Buena Vista Police Department**  
Address **306 Park Ave., Buena Vista, VA 24416**  
DFS Technician **Heather Stanton** License No. **18910**

☐ Laboratory

☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) **742** Reference Barometer (mm HG) **743**  
Reference Barometer(RB)Serial # **009113** RB Calibration Due **4/28/2015**

**Measurement Assurance Check**

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>	0.296
0.300	0.293	0.284	0.301	<b>Sample 2</b>	0.296
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>	0.296
0		0.296	0.296		

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>	0.098
0.100	0.098	0.095	0.101	<b>Sample 2</b>	0.098
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>	0.098
0		0.098	0.098		


Dry gas standard Lot No. (with tank no.) **AG404902-10**

☐ Replaced dry gas standard (+O-ring)


☐ Installed at Location

☒ Removed to DFS-Central

Supplies  
Mouthpieces  
Certificates of Analysis  
Operator Worksheet  
Other:



Notes:



Instrument Serial Number

010620

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	


Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section


DFS Technician



Date

4/9/15

Issuing Analyst



Date

4/13/15




COMMONWEALTH OF VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS  
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

HMS  
Jed

NAME OF ACCUSED <b>INSTRUMENT, TEST,</b>	NAME OF COURT <b>DFS</b>
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY <b>STANTON, HEATHER, M</b>		AGENCY <b>DFS Central Lab</b>
DFS LICENSE NUMBER <b>18910</b>	LICENSE EXPIRES <b>10/01/2016</b>	DATE TEST CONDUCTED <b>04/08/2015</b>
TEST EQUIPMENT NUMBER <b>010620</b>		

10:24 EDT

RESULTS: TIME SAMPLE TAKEN \_\_\_\_\_  
0.00

SAMPLE'S ALCOHOL CONTENT \_\_\_\_\_ GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS \_\_\_\_\_ DAY OF \_\_\_\_\_, 20 \_\_\_\_\_.

BREATH TEST OPERATOR
----------------------

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS \_\_\_\_\_  
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS \_\_\_\_\_  
OPERATOR'S SIGNATURE



# IntoxNet MIS Report

Report Generated 09 Apr 2015 at 08:59

*Handwritten initials: HMO and FUL*

## Test Results

Instrument Serial Number 010620

Test # 000903 Subject Test

Test Location 1 Department of  
Test Date 08 Apr 2015

Test Location 2 Forensic Science  
Test Time 10:18  
Remote/Local Local

Test Location 3  
System Check Passed

Operator's Last Name STANTON  
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 08 Apr 2015 End Time 10:26

Result Time 10:24

Result Date 08 Apr 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:19

Data Type BLK

Sample Value 0.000

Sample Time 10:19

Data Type CHK

Sample Value 0.098

Sample Time 10:20

Data Type BLK

Sample Value 0.000

Sample Time 10:21

Data Type SUBJ

Sample Value 0.000

Sample Time 10:22

Data Type BLK

Sample Value 0.000

Sample Time 10:23

Data Type SUBJ

Sample Value 0.000

Sample Time 10:24

Data Type BLK

Sample Value 0.000

Sample Time 10:25

Standard Type Dry Gas Std

Standard Value 0.098

Standard Lot Number AG404902-10

Standard Expiration Date 02/18/2016

Tank Pressure 992

Barometric Pressure 742 mmHg

Blow Sample Number 1 Blow Duration 3.30 sec

Blow Volume 1769 cc End-of-Blow Time 10:22

Blow Sample Number 2 Blow Duration 3.34 sec

Blow Volume 1760 cc End-of-Blow Time 10:24

Tamper Evident Stamp ec05ff51

Test Status Code 0

Test Status Success

THIS  
SID

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010620      Test Number: 904

Test Date: 04/08/2015      Test Time: 10:27 EDT

Dry Gas Target: 0.098

Lot Number: AG404902-10      Exp Date: 02/18/2016

Tank Pressure: 987 psi      Barometric Pressure: 741 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:27
CHK	0.098	10:28
BLK	0.000	10:30
CHK	0.098	10:30
BLK	0.000	10:32
CHK	0.098	10:32

Test Status: *Success*

Calibration CRC: 0C2FEE1E

HMUS  
Jrd

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010620      Test Number: 905

Test Date: 04/08/2015      Test Time: 10:39 EDT

Dry Gas Target: 0.293

Lot Number: AG506501-09      Exp Date: 03/06/2017

Tank Pressure: 365 psi      Barometric Pressure: 741 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:39
CHK	0.296	10:40
BLK	0.000	10:42
CHK	0.296	10:42
BLK	0.000	10:44
CHK	0.296	10:45

Test Status: *Success*

Calibration CRC: 0C2FEE1E